

C#001962

Statim File

Begin cycle # 001962

Date	Time
Jul 27 2018	9:40

Local Standard Time

STATIM 2000 S201R614
SN 100212G00016
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001962
10:48 07/27/2018
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:33
PRESSURIZING 1:38
STERILIZING 2:42
136.4 C 220kPa 2:42
136.2 C 224kPa 3:12
135.8 C 221kPa 3:42
136.5 C 227kPa 4:12
136.2 C 225kPa 4:42
136.5 C 226kPa 5:12
135.4 C 215kPa 5:42
136.4 C 227kPa 6:12
VENTING 6:12
AIR DRYING 7:03
CYCLE COMPLETE 34:33

Digital Key#
E8C6A7F2C402
